

测定范围阶段 (①~④自动切换)	①0~199 μ S/cm (1 μ S/cm)、②0.20~1.99ms/cm (0.01ms/cm)、 ③2.0~19.9ms/cm (0.1ms/cm)、④20~199ms/cm (1ms/cm) (闪烁显示(参考值))
测定精度	2%FS \pm 1 digit,但是当为10ms/cm以上时,测定精度为 3%FS \pm 1 digit 0~1.1%
食盐浓度换算显示范围	0~1.1%
显示范围(阶段)	①0.00~0.99% (0.01%)、②1.0~9.9% (0.1%)、 ③10~15% (1%) (②, ③/闪烁显示(参考值))